



Test Report

[2020] WSZ FHL NO.6100

Page 1 of 6

Product name	Filtering half mask	Specification	KQ-XY
		Brand	—
Client/Add/Tel	SUZHOU CARNATION MEDICAL APPARATUS CO.,LTD/NO. 3, FEIXIANG RD, KOREA INDUSTRIAL PARK, FENGHUANG TOWN, ZHANGJIAGANG CITY, JIANGSU PROVINCE, CHINA /13914912118		
Manufacturer/Add/Tel	SUZHOU CARNATION MEDICAL APPARATUS CO.,LTD/NO. 3, FEIXIANG RD, KOREA INDUSTRIAL PARK, FENGHUANG TOWN, ZHANGJIAGANG CITY, JIANGSU PROVINCE, CHINA /13914912118		
Sample grade	FFP2	Sample number	GW6100-2020
Sample quantity	110 pcs	Receiving date of sample	25/05/2020
Test type	Entrusted inspection	Article number/Batch number/Style number	—
Test date	26/05/2020-04/06/2020	Testing sites	Testing room
Sample state	Meeting the requirements of testing	Sample description	—
Test standard(s)	EN 149:2001+A1:2009 Respiratory protective devices-Filtering half masks to protect against particles- Requirements,testing, marking		
Test items	Visual inspection, practical performance, finish of parts, compatibility with skin, flammability, carbon dioxide content of the inhalation air, material, head harness, field of vision, penetration of filter material, breathing resistance, total inward leakage		
Test conclusion	<p>The sample upon testing, the test items meet the requirements of the EN 149:2001+A1:2009 standard. The detail of test results see on Pages 2-6.</p> <p style="text-align: center;">  Issue date: 05/06/2020 </p>		
Note	For the entrusted sample test, the technical responsibilities are undertaken for the test results of the supplied samples only.		

Approver:

[Signature]

Reviewer:

[Signature]

Chief Tester

[Signature]